



**PERFORMANCE WORK STATEMENT (PWS)
DEPARTMENT OF VETERANS AFFAIRS**

**Veterans Health Administration (VHA)
VA Center for Innovations (VACI)**

**Advanced Environmental Controls for Veterans
with Severe Disabilities**

Date: July 14, 2016

PWS Version Number: 1.2

**Advanced Environmental Controls for Veterans
with Severe Disabilities**

Contents

1.0	BACKGROUND.....	3
2.0	APPLICABLE DOCUMENTS.....	3
3.0	SCOPE OF WORK.....	6
4.0	PERFORMANCE DETAILS.....	6
4.1	PERFORMANCE PERIOD.....	6
4.2	PLACE OF PERFORMANCE.....	7
4.3	TRAVEL.....	7
5.0	SPECIFIC TASKS AND DELIVERABLES.....	8
5.1	PROJECT MANAGEMENT.....	8
5.1.1	PROJECT KICK-OFF MEETING.....	8
5.1.2	CONTRACTOR PROJECT MANAGEMENT PLAN.....	8
5.1.3	REPORTING REQUIREMENTS.....	9
5.2	SITE KICKOFF MEETING.....	9
5.3	ECU.....	10
5.4	ECU INSTALLATION.....	13
5.5	ECU CONFIGURATION.....	14
5.6	ACCEPTANCE AND TESTING.....	14
5.7	CLINICIAN TRAINING.....	15
5.8	TECHNICIAN TRAINING.....	15
5.9	OPTIONAL TASKS.....	16
5.9.1	TELEVISIONS AND ACCEPTANCE REPORT.....	16
5.9.2	ROLLING CARTS.....	16
5.9.3	BED COMPONENT KIT.....	17
5.9.4	ADDITIONAL TRAINING SESSIONS (CLINICAL AND TECHNICAL).....	17
6.0	GENERAL REQUIREMENTS.....	18
6.1	INSTALLATION OF DEVICES AND CUSTOM MOUNTING.....	18
6.2	ENTERPRISE AND IT FRAMEWORK.....	18
6.3	SECURITY AND PRIVACY REQUIREMENTS.....	20
6.3.1	POSITION/TASK RISK DESIGNATION LEVEL(S).....	20
6.3.2	CONTRACTOR PERSONNEL SECURITY REQUIREMENTS.....	22
6.4	METHOD AND DISTRIBUTION OF DELIVERABLES.....	24
6.5	PERFORMANCE METRICS.....	24
6.6	FACILITY/RESOURCE PROVISIONS.....	25
6.7	GOVERNMENT FURNISHED PROPERTY.....	25
6.8	SHIPMENT OF HARDWARE OR EQUIPMENT.....	25
	ADDENDUM A – ADDITIONAL VA REQUIREMENTS, CONSOLIDATED.....	25
	ADDENDUM B – VA INFORMATION AND INFORMATION SYSTEM SECURITY/PRIVACY LANGUAGE.....	33

Advanced Environmental Controls for Veterans with Severe Disabilities

1.0 BACKGROUND

The mission of the Department of Veterans Affairs (VA), VA Center for Innovations (VACI), and the Veterans Health Administration (VHA) Innovation Program is to field, fund, and foster ideas that benefit Veteran's health. The program allows mission critical healthcare innovations to emerge from the field and industry. These ideas, along with those from leadership, are matured through a collaborative constructive review by communities of interest, and are piloted in a safe harbor for innovation. Innovations that are proven and vetted from business and technical perspectives will have a pathway for organizational acceptance and diffusion.

This requirement is for the purchase and installation of Environmental Control Units (ECUs) for spinal cord injury patients, as well as to provide and install individual patient televisions with customized dual mounting.

Spinal cord injury patients with limited functionality do not have the capability to control their surroundings. The ECUs will provide an increased quality of patient care as well as patient safety.

2.0 APPLICABLE DOCUMENTS

In the performance of the tasks associated with this Performance Work Statement, the Contractor shall comply with the following:

1. 44 U.S.C. § 3541, "Federal Information Security Management Act (FISMA) of 2002"
2. Federal Information Processing Standards (FIPS) Publication 140-2, "Security Requirements For Cryptographic Modules"
3. FIPS Pub 201-2, "Personal Identity Verification of Federal Employees and Contractors," August 2013
4. 10 U.S.C. § 2224, "Defense Information Assurance Program"
5. Carnegie Mellon Software Engineering Institute, Capability Maturity Model® Integration for Development (CMMI-DEV), Version 1.3 November 2010; and Carnegie Mellon Software Engineering Institute, Capability Maturity Model® Integration for Acquisition (CMMI-ACQ), Version 1.3 November 2010
6. 5 U.S.C. § 552a, as amended, "The Privacy Act of 1974"
7. 42 U.S.C. § 2000d "Title VI of the Civil Rights Act of 1964"
8. VA Directive 0710, "Personnel Suitability and Security Program," June 4, 2010, <http://www.va.gov/vapubs/>
9. VA Handbook 0710, Personnel Suitability and Security Program, September 10, 2004, <http://www.va.gov/vapubs>
10. VA Directive and Handbook 6102, "Internet/Intranet Services," July 15, 2008
11. 36 C.F.R. Part 1194 "Electronic and Information Technology Accessibility Standards," July 1, 2003
12. Office of Management and Budget (OMB) Circular A-130, "Management of Federal Information Resources," November 28, 2000
13. 32 C.F.R. Part 199, "Civilian Health and Medical Program of the Uniformed Services (CHAMPUS)"

**Advanced Environmental Controls for Veterans
with Severe Disabilities**

14. An Introductory Resource Guide for Implementing the Health Insurance Portability and Accountability Act (HIPAA) Security Rule, October 2008
15. Sections 504 and 508 of the Rehabilitation Act (29 U.S.C. § 794d), as amended by the Workforce Investment Act of 1998 (P.L. 105-220), August 7, 1998
16. Homeland Security Presidential Directive (12) (HSPD-12), August 27, 2004
17. VA Directive 6500, "Managing Information Security Risk: VA Information Security Program," September 20, 2012
18. VA Handbook 6500, "Risk Management Framework for VA Information Systems – Tier 3: VA Information Security Program," March 10, 2015
19. VA Handbook 6500.1, "Electronic Media Sanitization," November 03, 2008
20. VA Handbook 6500.2, "Management of Breaches Involving Sensitive Personal Information (SPI)", October, 28, 2015
21. VA Handbook 6500.3, "Assessment, Authorization, And Continuous Monitoring Of VA Information Systems," February 3, 2014
22. VA Handbook 6500.5, "Incorporating Security and Privacy in System Development Lifecycle", March 22, 2010
23. VA Handbook 6500.6, "Contract Security," March 12, 2010
24. VA Handbook 6500.8, "Information System Contingency Planning", April 6, 2011
25. Project Management Accountability System (PMAS) portal (reference <https://www.voa.va.gov/pmas/>)
26. OI&T ProPath Process Methodology (reference process maps at <http://www.va.gov/PROPATH/Maps.asp> and templates at <http://www.va.gov/PROPATH/Templates.asp>)
27. One-VA Technical Reference Model (TRM) (reference at <http://www.va.gov/trm/TRMHomePage.asp>)
28. National Institute Standards and Technology (NIST) Special Publications (SP)
29. VA Directive 6508, Implementation of Privacy Threshold Analysis and Privacy Impact Assessment, October 15, 2014
30. VA Directive 6300, Records and Information Management, February 26, 2009
31. VA Handbook, 6300.1, Records Management Procedures, March 24, 2010
32. OMB Memorandum, "Transition to IPv6", September 28, 2010
33. VA Directive 0735, Homeland Security Presidential Directive 12 (HSPD-12) Program, February 17, 2011
34. VA Handbook 0735, Homeland Security Presidential Directive 12 (HSPD-12) Program, March 20, 2014
35. OMB Memorandum M-06-18, Acquisition of Products and Services for Implementation of HSPD-12, June 30, 2006
36. OMB Memorandum 05-24, Implementation of Homeland Security Presidential Directive (HSPD) 12 – Policy for a Common Identification Standard for Federal Employees and Contractors, August 5, 2005
37. OMB memorandum M-11-11, "Continued Implementation of Homeland Security Presidential Directive (HSPD) 12 – Policy for a Common

Advanced Environmental Controls for Veterans with Severe Disabilities

- Identification Standard for Federal Employees and Contractors, February 3, 2011
38. OMB Memorandum, Guidance for Homeland Security Presidential Directive (HSPD) 12 Implementation, May 23, 2008
 39. Federal Identity, Credential, and Access Management (FICAM) Roadmap and Implementation Guidance, December 2, 2011
 40. NIST SP 800-116, A Recommendation for the Use of Personal Identity Verification (PIV) Credentials in Physical Access Control Systems, November 20, 2008
 41. OMB Memorandum M-07-16, Safeguarding Against and Responding to the Breach of Personally Identifiable Information, May 22, 2007
 42. NIST SP 800-63-2, Electronic Authentication Guideline, August 2013
 43. Draft NIST Special Publication 800-157, Guidelines for Derived PIV Credentials, March 2014
 44. NIST Special Publication 800-164, Guidelines on Hardware-Rooted Security in Mobile Devices (Draft), October 2012
 45. Draft National Institute of Standards and Technology Interagency Report (NISTIR) 7981 Mobile, PIV, and Authentication, March 2014
 46. VA Memorandum, VAIQ #7100147, Continued Implementation of Homeland Security Presidential Directive 12 (HSPD-12), April 29, 2011 (reference <https://www.voa.va.gov/documentlistpublic.aspx?NodeID=514>)
 47. VA Memorandum, VAIQ # 7011145, VA Identity Management Policy, June 28, 2010 (reference Enterprise Architecture Section, PIV/IAM (reference <https://www.voa.va.gov/documentlistpublic.aspx?NodeID=514>))
 48. IAM Identity Management Business Requirements Guidance document, May 2013, (reference Enterprise Architecture Section, PIV/IAM (reference <https://www.voa.va.gov/documentlistpublic.aspx?NodeID=514>))
 49. Trusted Internet Connections (TIC) Reference Architecture Document, Version 2.0, Federal Interagency Technical Reference Architectures, Department of Homeland Security, October 1, 2013, https://www.fedramp.gov/files/2015/04/TIC_Ref_Arch_v2-0_2013.pdf
 50. OMB Memorandum M-08-05, "Implementation of Trusted Internet Connections (TIC)", November 20, 2007
 51. OMB Memorandum M-08-23, Securing the Federal Government's Domain Name System Infrastructure, August 22, 2008
 52. VA Memorandum, VAIQ #7497987, Compliance – Electronic Product Environmental Assessment Tool (EPEAT) – IT Electronic Equipment, August 11, 2014 (reference Document Libraries, EPEAT/Green Purchasing Section, <https://www.voa.va.gov/documentlistpublic.aspx?NodeID=552>)
 53. Sections 524 and 525 of the Energy Independence and Security Act of 2007, (Public Law 110–140), December 19, 2007
 54. Section 104 of the Energy Policy Act of 2005, (Public Law 109–58), August 8, 2005
 55. Executive Order 13693, "Planning for Federal Sustainability in the Next Decade", dated March 19, 2015

Advanced Environmental Controls for Veterans with Severe Disabilities

56. Executive Order 13221, "Energy-Efficient Standby Power Devices," August 2, 2001
57. VA Directive 0058, "VA Green Purchasing Program", July 19, 2013
58. VA Handbook 0058, "VA Green Purchasing Program", July 19, 2013
59. Office of Information Security (OIS) VAIQ #7424808 Memorandum, "Remote Access", January 15, 2014,
<https://www.voa.va.gov/DocumentListPublic.aspx?NodeId=28>
60. Clinger-Cohen Act of 1996, 40 U.S.C. §11101 and §11103
61. VA Directive 6071, Project Management Accountability System (PMAS), February 20, 2013
62. VA Memorandum, "Implementation of Federal Personal Identity Verification (PIV) Credentials for Federal and Contractor Access to VA IT Systems", (VAIQ# 7614373) July 9, 2015,
<https://www.voa.va.gov/DocumentListPublic.aspx?NodeId=28>
63. VA Memorandum "Mandatory Use of PIV Multifactor Authentication to VA Information System" (VAIQ# 7613595), June 30, 2015,
<https://www.voa.va.gov/DocumentListPublic.aspx?NodeId=28>
64. VA Memorandum "Mandatory Use of PIV Multifactor Authentication for Users with Elevated Privileges" (VAIQ# 7613597), June 30, 2015;
<https://www.voa.va.gov/DocumentListPublic.aspx?NodeId=28>
65. "Veteran Focused Integration Process Guide 1.0", December, 2015,
https://vaww.oit.va.gov/wp-content/uploads/2016/01/VIP_Guide_1_0_v14.pdf
66. "VIP Release Process Guide", Version 1.0, December 2015,
<https://www.voa.va.gov/DocumentView.aspx?DocumentID=4411>
67. "POLARIS User Guide", Version 1.2, February 2016,
<https://www.voa.va.gov/DocumentView.aspx?DocumentID=4412>

3.0 SCOPE OF WORK

The Contractor shall provide all necessary services and materials to configure, install and test Environmental Control Units for spinal cord injury patient beds as specified by each facility herein.

4.0 PERFORMANCE DETAILS

4.1 PERFORMANCE PERIOD

The period of performance shall be 12-months from date of award with 4 optional tasks.

Any work at the Government site shall not take place on Federal holidays or weekends unless directed by the Contracting Officer (CO).

There are ten (10) Federal holidays set by law (USC Title 5 Section 6103) that VA follows:

**Advanced Environmental Controls for Veterans
with Severe Disabilities**

Under current definitions, four are set by date:

New Year's Day	January 1
Independence Day	July 4
Veterans Day	November 11
Christmas Day	December 25

If any of the above falls on a Saturday, then Friday shall be observed as a holiday. Similarly, if one falls on a Sunday, then Monday shall be observed as a holiday.

The other six are set by a day of the week and month:

Martin Luther King's Birthday	Third Monday in January
Washington's Birthday	Third Monday in February
Memorial Day	Last Monday in May
Labor Day	First Monday in September
Columbus Day	Second Monday in October
Thanksgiving	Fourth Thursday in November

4.2 PLACE OF PERFORMANCE

Tasks under this PWS shall be performed in the below listed VA facilities:

1. Minneapolis Veteran Affairs Medical Center (VAMC), Minneapolis, MN
2. VA Palo Alto Healthcare System, Palo Alto, CA
3. VA Puget Sound Healthcare System, Seattle, WA
4. VA New Mexico Health Care System, Albuquerque, NM
5. Clement J. Zablocki VAMC, Milwaukee, WI
6. Saint Louis Health Care System, St. Louis, MO
7. VA San Diego Health Care System, San Diego, CA
8. Hampton VAMC, Hampton, VA

4.3 TRAVEL

The Government anticipates travel under this effort to perform the tasks associated with the effort, as well as to attend program-related meetings or conferences throughout the PoP. Include all estimated travel costs in your firm-fixed price line items. The below listed travel accounts only for the base year tasks. Any additional travel required for the optional task 5.9.4 shall be priced separately. These costs will not be directly reimbursed by the Government.

LOCATION	TRIPS	DURATION DAYS
Minneapolis VAMC, Minneapolis, MN	6	5
VA Palo Alto Healthcare System, Palo Alto, CA	9	5

**Advanced Environmental Controls for Veterans
with Severe Disabilities**

VA Puget Sound Healthcare System, Seattle, WA	6	5
VA New Mexico Health Care System, Albuquerque, NM	6	5
Clement J. Zablocki VAMC, Milwaukee, WI	8	5
Saint Louis Health Care System, St. Louis, MO	8	5
VA San Diego Health Care System, San Diego, CA	6	5
Hampton VAMC, Hampton, VA	9	5

5.0 SPECIFIC TASKS AND DELIVERABLES

The Contractor shall perform the following:

5.1 PROJECT MANAGEMENT

5.1.1 PROJECT KICK-OFF MEETING

The Contractor shall hold a Project Kick-Off Meeting with an Integrated Project Team (IPT) consisting of all key stakeholders including the Contracting Officer’s Representative (COR) and CO within 10 calendar days of contract award. At the Project Kick-Off Meeting, the Contractor shall present the details of its intended approach, work plan, Contractor Project Management Plan (CPMP) detailed in section 5.1.2, and project schedule, including deliverable dates, for review and approval by the VA Program Manager (PM) and COR. The Contractor shall provide a Project Kick-Off Meeting Agenda, Project Kick-Off Meeting Briefing and Project Kick-Off Meeting Minutes.

Deliverable:

- A. Project Kick-Off Meeting Briefing

5.1.2 CONTRACTOR PROJECT MANAGEMENT PLAN

The Contractor shall deliver a Contractor Project Management Plan (CPMP) that lays out the Contractor’s approach, timeline and tools to be used in execution of the contract. The CPMP should take the form of both a narrative and graphic format that displays the schedule, milestones, risks and resource support. The CPMP shall also include how the Contractor shall coordinate and execute planned, routine, and ad hoc data collection reporting requests as identified within the PWS. The initial baseline

Advanced Environmental Controls for Veterans with Severe Disabilities

CPMP shall be concurred upon and updated in accordance with Section B of the contract. The Contractor shall update and maintain the VA PM approved CPMP throughout the PoP.

Deliverable:

- A. Contractor Project Management Plan

5.1.3 REPORTING REQUIREMENTS

The Contractor shall provide the COR with Monthly Progress Reports in electronic form in Microsoft Word and Project formats. The report shall include detailed instructions/explanations for each required data element, to ensure that data is accurate and consistent. These reports shall reflect data as of the last day of the preceding Month.

The Monthly Progress Reports shall cover all work completed during the reporting period and work planned for the subsequent reporting period. The report shall also identify any problems that arose and a description of how the problems were resolved. If problems have not been completely resolved, the Contractor shall provide an explanation including their plan and timeframe for resolving the issue. The report shall also include an itemized list of all Electronic and Information Technology (EIT) deliverables and their current Section 508 conformance status. The Contractor shall monitor performance against the CPMP and report any deviations. It is expected that the Contractor will keep in communication with VA accordingly so that issues that arise are transparent to both parties to prevent escalation of outstanding issues.

Deliverable:

- A. Monthly Progress Report

5.2 SITE KICKOFF MEETING

The Contractor shall participate in a Site Kick-off Meeting (SKM) at each of the 8 VA facilities detailed in section 4.2 to be held 30 days after contract award. The purpose of the SKM for each individual site is to ensure that there is an understanding between VA and the Contractor of the work to be executed at the site, the project approach, major actions, the schedule for key milestones and deliverables, and what is needed from VA and the Contractor to ensure project success. For each individual site, the Contractor shall provide an SKM Briefing which shall review task coordination details, schedules and topics, deliverable review and issue resolution processes and preferred site communication mechanisms specific to that site. The Contractor shall provide a Site Kick-Off Meeting Agenda and Site Kick-Off Meeting Minutes for each of the 8 Site Kick-Offs. The Contractor shall also provide a Site Kick-Off Briefing for each of the 8 Site Kick-Offs.

Advanced Environmental Controls for Veterans with Severe Disabilities

Deliverables:

- A. Site Kick-Off Meeting Briefing for Minneapolis VAMC
- B. Site Kick-Off Meeting Briefing for VA Palo Alto Healthcare System
- C. Site Kick-Off Meeting Briefing for VA Puget Sound Healthcare System
- D. Site Kick-Off Meeting Briefing for VA New Mexico Health Care System
- E. Site Kick-Off Meeting Briefing for Clement J. Zablocki VAMC
- F. Site Kick-Off Meeting Briefing for Saint Louis Healthcare System
- G. Site Kick-Off Meeting Briefing for VA San Diego Healthcare System
- H. Site Kick-Off Meeting Briefing for Hampton VAMC

5.3 ECU

The Contractor shall provide an Environmental Control Unit (ECU) that has the following capabilities. Minimum capabilities should resemble the Autonomie Environmental Care Unit or equivalent device with salient characteristics as follows:

1. ECU Core Unit:
 - a. At least 4 GB RAM
 - b. At least 80 GB Solid State Drive
 - c. Windows 7 or latest version operating system
 - d. Touch screen
 - e. Can be programmed to perform no fewer than 150 functions
 - f. Full on-board programming, no tools or hardware required
 - g. Built-in infrared capture
 - h. Macros (scenes) fully editable with up to 20 functions per macro
 - i. User-adjustable volume, brightness, and scan rate
 - j. Custom template design functions
 - k. Minimum of 12.1" LED touchscreen
 - l. Minimum of 6-7 hour battery life
 - m. Seamless File Backup & Management
 - n. One Button Press Design for Remote Support Team Tools
 - o. Remote Assistance
 - p. Wi-Fi / Dual Mode Bluetooth 4.0 capabilities
 - q. 4 USB Ports (minimum 2.0)
 - s. 1 Headphone/1 Microphone jack 1/8"
 - t. Integrated Universal Infrared Remote (IR) Capabilities with learning window and multiple IR emitters
 - u. Integrated Z-Wave capability for environmental control utilization
 - v. Dual switch ports for ancillary access equipment
 - w. Assistive Technology software designed for application integration and multiple access methodologies
 - x. Ability to mount the device to a bed as well as a wheelchair if necessary
 - y. Minimum of 120 square foot coverage range
2. Voice Recognition and Control
 - a. Input

Advanced Environmental Controls for Veterans with Severe Disabilities

- 1) Speech
- 2) Dual switch or single switch with scanning
- 3) At least 96 unique and different words available with multiple command use
- 4) Internal microphone
- 5) External keys for caregiver use
- b. Output
 - 1) Internal Loudspeaker
 - 2) Auditory – provide voice feedback on menu commands that will acknowledge confirmation on a patient's voice command.
 - 3) Additional ECU Accessories
3. Head Controlled Mouse
 - a. Operating Wave Band: Near Infrared
 - b. Capable of Wireless Operation
 - c. Power Consumption: 1 Watt, minimum
 - d. Field of view: At least 45-degrees by 60-degrees
 - e. Standard Target: At least 0.25 inch (6.5mm) Diameter
 - f. Measurement Rate: 45 Hz
 - g. Measurement Latency: 10 msec.
4. Communication
6. Joints for articulating arms shall allow for a minimum of 180 degree of rotation on any axis.
7. Base for mounting arms shall have the ability to transfer locations
8. ECU shall be capable of being controlled by eye gaze with the following characteristics:
 - a. Tracking – Hybrid infrared video eye & binocular & monocular tracking
 - b. Working volume – minimum of 300x200x200mm³ (WxHxD)
 - c. Accuracy, static - 0.5 degree
 - d. Accuracy, over full working volume – 1 degree
9. ECU shall be capable of being controlled by sip n' puff tubing and straws that are commercially available
10. ECU shall have single and dual switch scanning
11. ECU shall provide access and interfaces to utilize computer games, E-books, ability to write documents, internet access and social video networking applications
 - a. Wall bracket that is compatible with the dual mounting for televisions
12. ECU shall interface with nurse call systems.
13. All cabling required to interface with the hospital equipment shall be included, including:
 - a. Nurse Call cables with ¼" male jacks to 1/4" jacks
 - b. Switch Cables from ECU device to bed unit
 - c. Additional bed cables will be determined by bed type. Bed type will be verified in field.
14. Compatible with the following bed types at selected facilities:

Advanced Environmental Controls for Veterans with Severe Disabilities

- a. Stryker: GoBed+, Gobed 2, Secure 2, Patriot, Hill Rom Care Assist Bed Model, Hill-Rom Advanta 350, Hill-Rom Flexicare II, Hill-Rom Careassist, Hill-Rom Versacare, Hill Rom Clinitron bed, Sizewise Navigator and Hill-Rom Progressa
15. Mounting for the sip-n-puff straws, microphones, speakers, and other accessories must be included for the beds.
 16. Augmentative communication capabilities to include word prediction software must also be included.
 17. Ability to delete user settings and saved documents in between patients must be included.
 18. Break away cable with a single connection for control by patient care providers (bed)
 19. Cable management to include cable coverings and protection (bed)
 20. Software interface for ECU capable of being minimized for full PC features by administrative staff
 21. Dual mount with minimum capabilities that shall resemble the PDi Communication System Mod PD168-046 Dual Monitor Mount or equivalent device with salient characteristics as follows:
 1. The mounting for the televisions shall be designed to be dual mounting and support both the weight of the television as well as the weight of the environmental control unit.
 2. The mounting bracket shall be able to rotate approximately 300 degrees and have +/- 15 degrees of tilt towards and away from the patient in order to accommodate special needs of patients.
 3. The arm for the mounting shall have a channel for nurse call.
 4. An external channel shall be supplied on the outside of the arm for the cabling for the environmental control unit.
 5. Installation at every patient bed in the spinal cord unit shall minimize disruption to patient care.

The Contractor shall warrant that the products they install shall be free from defects in materials and workmanship for a period of no less than two years from the date of acceptance. If any such product proves defective during the warranty period the vendor will repair the defective product at no charge to the Government for parts and labor, or a replacement in exchange for the defective product.

At several VA Medical Centers, the ECU system may be installed in concert with other manufacturer's hardware and solutions ("other VA vendors"). In the event of a mechanical or technical issue for these instances, the Contractor shall conduct a root cause analysis to determine what the issue is and notify the Government which solution caused the issue, and the party responsible for causing the issue remediate and provide maintenance as described herein. Because other VA vendors' products and services within this contract's place of performance may result in reduced functionality or impaired performance of Contractor's solution, the Government will notify Contractor reasonably in advance of any updates, patches, or changes to the other VA vendors' technology or related services that may impact the delivery of supplies and services,

Advanced Environmental Controls for Veterans with Severe Disabilities

delivery schedule or supplier's warranty obligations. Contractor can make no representations or warranties of any kind regarding the products or services of other VA vendors (not subcontracted through Contractor), including, without limitation, the compatibility of such products or services with Contractor's solution. Contractor will be available to further assist in resolving any such issues, however, since such assistance may be considered out of scope, Contractor will seek prior approval from the Contracting Officer in accordance with the terms and conditions of the contract.

If the parties agree that any proposed resolution materially or significantly changes Contract's scope of work or pricing, the Contractor shall not perform such work. These changes will be handled through mutual agreement of the parties in accordance with the commercial changes clause found at FAR 52.212-4(c). Contractor will notify the Government in writing if a proposed solution impacting its delivery of supplies/services and/or technology cannot be further developed, designed, or adapted to be compatible with other VA vendors' work, or if a proposed solution limits or voids supplier's warranty.

The Contractor shall provide ECU's with the above requirements to the below VA facilities:

- 30 ECUs at Minnesota VAMC, Minneapolis, MN
- 43 ECUs at VA Palo Alto Healthcare System, Palo Alto, CA
- 30 ECUs at VA Puget Sound Healthcare System, Seattle, WA
- 30 ECUs at VA New Mexico Health Care System, Albuquerque, NM
- 38 ECUs at Clement J. Zablocki VAMC, Milwaukee WI
- 38 ECUs at Saint Louis Health Care System, St. Louis, MO
- 30 ECUs at VA San Diego Health Care System, San Diego, CA
- 45 ECUs at Hampton VAMC, Hampton, VA

Deliverables:

- A. ECUs at Minnesota VAMC, Minneapolis, MN
- B. ECUs at VA Palo Alto Healthcare System, Palo Alto, CA
- C. ECUs at VA Puget Sound Healthcare System, Seattle, WA
- D. ECUs at VA New Mexico Health Care System, Albuquerque, NM
- E. ECUs at Clement J. Zablocki VAMC, Milwaukee WI
- F. ECUs at Saint Louis Health Care System, St. Louis, MO
- G. ECUs at VA San Diego Health Care System, San Diego, CA
- H. ECUs at Hampton VAMC, Hampton, VA

5.4 ECU INSTALLATION

The Contractor shall install ECU's in patient rooms. In addition to the installation of the environmental control units, the arms and accessories such as the sip n puff straw, microphone, etc. shall also be installed on patient beds. The entire system shall be installed and functional at the conclusion of the installation to include accessories. This will be coordinated to occur with minimal disruption to the patients.

Advanced Environmental Controls for Veterans with Severe Disabilities

5.5 ECU CONFIGURATION

The Contractor shall configure of all units including the environmental control system units along with the command center, phone with privacy speaker port, microphone input, microphone, direct IR input and wall mount bracket; Command Center Interface with IR relay, 4 way bed control, IR receiver and Nurse call. VA has an additional requirement for a "break away cable" on the wall panel due to safety concerns. The Contractor shall place the wall components on a pre-assembled wall mounted panel that shall contain the telephone and command center. From this panel will be a cable leading to the bed called a "break away cable". This cable will contain the bed control, microphone, and sip and puff switch wires. The cable will be in two pieces, with a disconnect capability within 2 to 4 feet from the bed to allow the nursing staff or others to detach the bed from the environmental control before moving the bed.

The Contractor shall provide an ECU Configuration Report for each ECU at each VA facility that details the installation and configuration of each ECU as it is configured.

Deliverable:

- A. ECU Configuration and Reports at Minnesota VAMC
- B. ECU Configuration and Reports at VA Palo Alto Healthcare System
- C. ECU Configuration and Reports at VA Puget Sound Healthcare System
- D. ECU Configuration and Reports at VA New Mexico Health Care System
- E. ECU Configuration and Reports at Clement J. Zablocki VAMC
- F. ECU Configuration and Reports at Saint Louis Health Care System
- G. ECU Configuration and Reports at VA San Diego Health Care System
- H. ECU Configuration and Reports at Hampton VAMC

5.6 ACCEPTANCE AND TESTING

The Contractor shall demonstrate the successful operation of each ECU in each facility in the presence of a Government Test Witness. The Government will accept the units installed and configured upon successful completion of the above mentioned demonstration of their ability to perform each of the requirements detailed in section 5.3 above.

Deliverable:

- A. Customer Acceptance Report for 30 beds at Minnesota VAMC
- B. Customer Acceptance Report for 43 beds at VA Palo Alto Healthcare System
- C. Customer Acceptance Report for 30 beds at VA Puget Sound Healthcare System
- D. Customer Acceptance Report for 30 beds at VA New Mexico Health Care System
- E. Customer Acceptance Report for 38 beds at Clement J. Zablocki VAMC
- F. Customer Acceptance Report for 38 beds at Saint Louis Health Care System
- G. Customer Acceptance Report for 30 beds at VA San Diego Health Care System
- H. Customer Acceptance Report for 45 beds at Hampton VAMC

Advanced Environmental Controls for Veterans with Severe Disabilities

5.7 CLINICIAN TRAINING

The Contractor shall develop and deliver train-the-trainer training during the site acceptance period to clinicians on the use of the ECUs. The goal of the clinician training will be to allow them to train the patients in the operation of the ECU and bedside television. The training for clinicians shall include a review of the ECU Quick start Guide, User operating manuals, handouts, and hands-on training to cover all of the ECU functionality. The training sessions shall be on-site and shall be approximately four hours in duration. The Contractor shall provide all Clinician Training Materials which shall include a Quick Start Guide and User Manual. The Contractor shall also provide all handouts used during training.

The number of personnel to be trained at each facility is identified below.

1. Minneapolis VAMC, Minneapolis, MN – 13 Clinicians
2. VA Palo Alto Healthcare System, Palo Alto, CA – 13 Clinicians
3. VA Puget Sound Healthcare System, Seattle, WA – 13 Clinicians
4. VA New Mexico Health Care System, Albuquerque, NM – 13 Clinicians
5. Clement J. Zablocki VAMC, Milwaukee, WI – 13 Clinicians
6. Saint Louis Health Care System, St. Louis, MO – 13 Clinicians
7. VA San Diego Health Care System, San Diego, CA – 13 Clinicians
8. Hampton VAMC, Hampton, VA – 13 Clinicians

Deliverables:

- A. Clinician Training Material Minneapolis VAMC
- B. Clinician Training Material VA Palo Alto Healthcare System
- C. Clinician Training Material VA Puget Sound Healthcare System
- D. Clinician Training Material VA New Mexico Health Care System
- E. Clinician Training Material Clement J. Zablocki VAMC
- F. Clinician Training Material Saint Louis Health Care System, St. Louis, MO
- G. Clinician Training Material VA San Diego Health Care System
- H. Clinician Training Material VAMC, Hampton, VA

5.8 TECHNICIAN TRAINING

The Contractor shall develop and deliver a separate training session for site selected technician personnel on troubleshooting and maintaining the equipment during the site acceptance period. This Contractor shall provide Technician Training Material which shall include the Quick Start Guide, User Manual, Maintenance and Troubleshoot Guide. The Contractor shall also provide all of the handouts used during training. The Technician Training Material shall also contain the necessary schematics and part lists for the VA to maintain the equipment after the expiration of the Contractor's warranty period. The training session shall not exceed four hours in duration.

The number of personnel to be trained at each facility is identified below.

1. Minneapolis VAMC, Minneapolis, MN – 3 Technicians
2. VA Palo Alto Healthcare System, Palo Alto, CA – 3 Technicians
3. VA Puget Sound Healthcare System, Seattle, WA – 3 Technicians

Advanced Environmental Controls for Veterans with Severe Disabilities

4. VA New Mexico Health Care System, Albuquerque, NM – 3 Technicians
5. Clement J. Zablocki VAMC, Milwaukee, WI – 3 Technicians
6. Saint Louis Health Care System, St. Louis, MO – 3 Technicians
7. VA San Diego Health Care System, San Diego, CA – 3 Technicians
8. Hampton VAMC, Hampton, VA – 3 Technicians

Deliverables:

Technician Training Material Minneapolis VAMC
Technician Training Material VA Palo Alto Healthcare System
Technician Training Material VA Puget Sound Healthcare System
Technician Training Material VA New Mexico Health Care System
Technician Training Material Clement J. Zablocki VAMC
Technician Training Material Saint Louis Health Care System
Technician Training Material VA San Diego Health Care System
Technician Training Material Hampton VAMC, Hampton, VA

5.9 OPTIONAL TASKS

5.9.1 TELEVISIONS AND ACCEPTANCE REPORT

The Contractor shall provide televisions to locations specified in the base period as well as acceptance reports outlining Contractor's conformance to specifications listed below. This optional task can be exercised up to 284 times to provide VA with up to 284 televisions.

The televisions shall be:

1. Flat screen LED and have a minimum 12 inch and maximum 15 inch diagonal sized screen.
2. Resolution 1080p
3. Individual televisions that are mounted on a wall arm that extends for the television to be in front of the patient for use.

The Contractor shall provide all labor, supervision, and materials for the installation of new televisions.

Deliverables:

- A. Televisions and Acceptance Report for Televisions

5.9.2 ROLLING CARTS

In the event the regular bed is not available or is replaced with a temporary substitute, the Contractor shall provide a rolling cart in which the ECU components will be mounted. The cart must meet the following specifications listed below:

- a. Locking wheels
- b. A minimum size work surface of 500 square inches for mounting of flexible arm with microphone, sip n' puff, and components
- c. Weight capacity of 150 lbs

Advanced Environmental Controls for Veterans with Severe Disabilities

d. Height adjustable from 27.5" – 37.5"

This optional task can be exercised up to 80 times to provide VA with up to 80 rolling carts.

Deliverables:

A. Rolling Carts

5.9.3 BED COMPONENT KIT

The Contractor shall provide bed component kits to locations specified in the base period as well as acceptance reports outlining Contractor's conformance to specifications listed below.

The bed component kits shall include the following and meet the minimum capabilities as referenced:

1. Bed mounted flexible arm with microphone and sip n' puff (as specified in section 5.2 #7, #9, and #15)
2. Sip n' Puff Switch (as specified in section 5.2 #10, #13b, and #15)
3. Bed box with breakaway cable (as specified in section 5.2 #18 and #19)
4. Bed controller (as specified in section 5.2 #13d)

The Contractor shall provide all labor, supervision, and materials for the installation of bed component kits.

This optional task can be exercised up to 80 times to provide VA with up to 80 bed component kits.

Deliverables:

A. Bed Component Kits and Acceptance Report

5.9.4 ADDITIONAL TRAINING SESSIONS (CLINICAL AND TECHNICAL)

The Contractor shall deliver an additional train-the-trainer training to clinicians and technicians on the use of the Environmental Control Units. Additional training shall be consistent with requirements outlined in sections 5.7 and 5.8 and be all-inclusive of travel.

This optional task can be exercised up to 16 times to provide VA with up to 16 sessions.

Deliverables:

A. Clinician and Technical Training and Materials

Advanced Environmental Controls for Veterans with Severe Disabilities

6.0 GENERAL REQUIREMENTS

6.1 INSTALLATION OF DEVICES AND CUSTOM MOUNTING

A) The Contractor shall completely prepare site for removal and installation of equipment. The Contractor shall furnish all labor and materials and perform work as required by specifications.

B) All employees of the Contractor and their Subcontractors shall comply with VA security management program.

C) Restoration:

The Contractor shall clean up all work areas after completing work in VA facilities; repair any damages including removal and disposal of defective equipment. The Contractor shall immediately repair and/or replace all facilities and/or equipment damaged by the Contractor and/or their Subcontractors. All areas affected shall be restored to their original condition. The Contractor shall include a Damage Incident log in their monthly progress report submitted to the Government detailing any damages as well as their restoration activities and completion

D) Final Cleanup: Upon completion of the project, or as work progresses, The Contractor shall remove all installation debris that has been part of the installation.

6.2 ENTERPRISE AND IT FRAMEWORK

The Contractor shall support the VA enterprise management framework. In association with the framework, the Contractor shall comply with OI&T Technical Reference Model (One-VA TRM). One-VA TRM is one component within the overall Enterprise Architecture (EA) that establishes a common vocabulary and structure for describing the information technology used to develop, operate, and maintain enterprise applications. One-VA TRM includes the Standards Profile and Product List that collectively serves as a VA technology roadmap. Architecture, Strategy, and Design (ASD) has overall responsibility for the One-VA TRM.

The Contractor shall ensure Commercial Off-The-Shelf (COTS) product(s), software configuration and customization, and/or new software are PIV-enabled by accepting HSPD-12 PIV credentials using VA Enterprise Technical Architecture (ETA), http://www.ea.oit.va.gov/VA_EA/VAEA_TechnicalArchitecture.asp, and VA Identity and Access Management (IAM) approved enterprise design and integration patterns, http://www.techstrategies.oit.va.gov/enterprise_dp.asp. The Contractor shall ensure all Contractor delivered applications and systems are compliant with VA Identity Management Policy (VAIQ# 7011145), Continued Implementation of Homeland Security Presidential Directive 12 (VAIQ#7100147), and VA IAM enterprise identity management requirements (IAM Identity Management Business Requirements Guidance document), located at <https://www.voa.va.gov/documentlistpublic.aspx?NodeID=514>. The Contractor shall ensure all Contractor delivered applications and systems provide user authentication services compliant with NIST Special Publication 800-63, VA Handbook 6500 Appendix F, "VA System Security Controls", and VA IAM enterprise requirements for direct, assertion based authentication, and/or trust based authentication, as determined by the design and integration patterns. Direct authentication at a minimum must include Public Key Infrastructure (PKI) based authentication supportive of

Advanced Environmental Controls for Veterans with Severe Disabilities

Personal Identity Verification (PIV) and/or Common Access Card (CAC), as determined by the business need. Assertion based authentication must include a SAML implementation. Additional assertion implementations, besides the required SAML assertion, may be provided as long as they are compliant with NIST 800-63 guidelines. Trust based authentication must include authentication/account binding based on trusted HTTP headers. The Contractor solution shall conform to the specific Identity and Access Management PIV requirements are set forth in OMB Memoranda M-04-04 (<http://www.whitehouse.gov/sites/default/files/omb/memoranda/fy04/m04-04.pdf>), M-05-24 (<http://www.whitehouse.gov/sites/default/files/omb/memoranda/fy2005/m05-24.pdf>), M-11-11 (<http://www.whitehouse.gov/sites/default/files/omb/memoranda/2011/m11-11.pdf>), National Institute of Standards and Technology (NIST) Federal Information Processing Standard (FIPS) 201-2, and supporting NIST Special Publications.

The Contractor solution shall support the latest Internet Protocol Version 6 (IPv6) based upon the directive issued by the Office of Management and Budget (OMB) on September 28, 2010 (<https://cio.gov/wp-content/uploads/downloads/2012/09/Transition-to-IPv6.pdf>) & (<http://www.cybertelecom.org/dns/ipv6usg.htm>). IPv6 technology, in accordance with the USGv6: A Technical Infrastructure for USGv6 Adoption (<http://www.nist.gov/itl/antd/usgv6.cfm>) and the NIST SP 800 series applicable compliance (<http://csrc.nist.gov/publications/PubsSPs.html>), shall be included in all IT infrastructures, application designs, application development, operational systems and sub-systems, and their integration. All public/external facing servers and services (e.g. web, email, DNS, ISP services, etc.) shall support native IPv6 users, including all internal infrastructure and applications shall communicate using native IPv6 operations. Guidance and support of improved methodologies which ensure interoperability with legacy protocol and services, in addition to OMB/VA memoranda, can be found at <https://www.voa.va.gov/documentlistpublic.aspx?NodeID=282>.

The Contractor solution shall meet the requirements outlined in Office of Management and Budget Memorandum M08-05 mandating Trusted Internet Connections (TIC) (<http://www.whitehouse.gov/sites/default/files/omb/assets/omb/memoranda/fy2008/m08-05.pdf>), M08-23 mandating Domain Name System Security (NSSEC) (<http://www.whitehouse.gov/sites/default/files/omb/assets/omb/memoranda/fy2008/m08-23.pdf>), and shall comply with the Trusted Internet Connections (TIC) Reference Architecture Document, Version 2.0 (https://www.fedramp.gov/files/2015/04/TIC_Ref_Arch_v2-0_2013.pdf).

The Contractor IT end user solution that is developed for use on standard VA computers shall be compatible with and be supported on the standard VA operating system, currently Windows 7 (64bit), Internet Explorer 11 and Microsoft Office 2010. In preparation for the future VA standard configuration update, end user solutions shall also be compatible with Office 2013 and Windows 8.1. However, Office 2013 and Windows 8.1 are not the VA standard yet and are currently not approved for use on the VA Network, but are in-process for future approval by OI&T. Upon the release approval of Office 2013 and Windows 8.1 individually as the VA standard, Office 2013 and

Advanced Environmental Controls for Veterans with Severe Disabilities

Windows 8.1 will supersede Office 2010 and Windows 7 respectively. Applications delivered to the VA and intended to be deployed to Windows 7 workstations shall be delivered as a signed .msi package and updates shall be delivered in signed .msp file formats for easy deployment using System Center Configuration Manager (SCCM) VA's current desktop application deployment tool. Signing of the software code shall be through a vendor provided certificate that is trusted by the VA using a code signing authority such as Verizon/Cybertrust or Symantec/VeriSign. The Contractor shall also ensure and certify that their solution functions as expected when used from a standard VA computer, with non-admin, standard user rights that have been configured using the United States Government Configuration Baseline (USGCB) specific to the particular client operating system being used.

The Contractor shall support VA efforts in accordance with the Project Management Accountability System (PMAS) that mandates all new VA IT projects/programs use an incremental development approach, requiring frequent delivery milestones that deliver new capabilities for business sponsors to test and accept functionality. Implemented by the Assistant Secretary for IT, PMAS is a VA-wide initiative to better empower the OI&T Project Managers and teams to meet their mission: delivering world-class IT products that meet business needs on time and within budget.

The Contractor shall utilize ProPath, the OI&T-wide process management tool that assists in the execution of an IT project. It is a one-stop shop providing critical links to the formal approved processes, artifacts, and templates to assist project teams in facilitating their VIP and PMAS-compliant work.

6.3 SECURITY AND PRIVACY REQUIREMENTS

6.3.1 POSITION/TASK RISK DESIGNATION LEVEL(S)

Position Sensitivity	Background Investigation (in accordance with Department of Veterans Affairs 0710 Handbook, "Personnel Suitability and Security Program," Appendix A)
Low / Tier 1	Tier 1 / National Agency Check with Written Inquiries (NACI) A Tier 1/NACI is conducted by OPM and covers a 5-year period. It consists of a review of records contained in the OPM Security Investigations Index (SII) and the DOD Defense Central Investigations Index (DCII), Federal Bureau of Investigation (FBI) name check, FBI fingerprint check, and written inquiries to previous employers and references listed on the application for employment. In VA it is used for Non-sensitive or Low Risk positions.

**Advanced Environmental Controls for Veterans
with Severe Disabilities**

Position Sensitivity	Background Investigation (in accordance with Department of Veterans Affairs 0710 Handbook, "Personnel Suitability and Security Program," Appendix A)
Moderate / Tier 2	Tier 2 / Moderate Background Investigation (MBI) A Tier 2/MBI is conducted by OPM and covers a 5-year period. It consists of a review of National Agency Check (NAC) records [OPM Security Investigations Index (SII), DOD Defense Central Investigations Index (DCII), FBI name check, and a FBI fingerprint check], a credit report covering a period of 5 years, written inquiries to previous employers and references listed on the application for employment; an interview with the subject, law enforcement check; and a verification of the educational degree.
High / Tier 4	Tier 4 / Background Investigation (BI) A Tier 4/BI is conducted by OPM and covers a 10-year period. It consists of a review of National Agency Check (NAC) records [OPM Security Investigations Index (SII), DOD Defense Central Investigations Index (DCII), FBI name check, and a FBI fingerprint check report], a credit report covering a period of 10 years, written inquiries to previous employers and references listed on the application for employment; an interview with the subject, spouse, neighbors, supervisor, co-workers; court records, law enforcement check, and a verification of the educational degree.

The position sensitivity and the level of background investigation commensurate with the required level of access for the following tasks within the PWS are:

Position Sensitivity and Background Investigation Requirements by Task

Task Number	Tier1 / Low / NACI	Tier 2 / Moderate / MBI	Tier 4 / High / BI
5.1	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
5.2	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
5.3	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
5.4	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
5.5	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
5.6	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
5.7	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
5.8	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
5.9	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>

The Tasks identified above and the resulting Position Sensitivity and Background Investigation requirements identify, in effect, the Background Investigation requirements for Contractor individuals, based upon the tasks the particular Contractor individual will be working. The submitted Contractor Staff Roster must indicate the required

Advanced Environmental Controls for Veterans with Severe Disabilities

Background Investigation Level for each Contractor individual based upon the tasks the Contractor individual will be working, in accordance with their submitted proposal.

6.3.2 CONTRACTOR PERSONNEL SECURITY REQUIREMENTS

Contractor Responsibilities:

- a. The Contractor shall prescreen all personnel requiring access to the computer systems to ensure they maintain the appropriate Background Investigation, and are able to read, write, speak and understand the English language.
- b. The Contractor shall bear the expense of obtaining background investigations.
- c. Within 3 business days after award, the Contractor shall provide a roster of Contractor and Subcontractor employees to the COR to begin their background investigations in accordance with the ProPath template. The Contractor Staff Roster shall contain the Contractor's Full Name, Date of Birth, Place of Birth, individual background investigation level requirement (based upon Section 6.2 Tasks), etc. The Contractor shall submit full Social Security Numbers either within the Contractor Staff Roster or under separate cover to the COR. The Contractor Staff Roster shall be updated and provided to VA within 1 day of any changes in employee status, training certification completion status, Background Investigation level status, additions/removal of employees, etc. throughout the Period of Performance. The Contractor Staff Roster shall remain a historical document indicating all past information and the Contractor shall indicate in the Comment field, employees no longer supporting this contract. The preferred method to send the Contractor Staff Roster or Social Security Number is by encrypted e-mail. If unable to send encrypted e-mail, other methods which comply with FIPS 140-2 are to encrypt the file, use a secure fax, or use a traceable mail service.
- d. The Contractor should coordinate the location of the nearest VA fingerprinting office through the COR. Only electronic fingerprints are authorized.
- e. The Contractor shall ensure the following required forms are submitted to the COR within 5 days after contract award:
 - 1) For a Tier 1/Low Risk designation:
 - a) OF-306
 - b) DVA Memorandum – Electronic Fingerprints
 - 2) For Tier 2/Moderate or Tier 4/High Risk designation:
 - a) OF-306
 - b) VA Form 0710
 - c) DVA Memorandum – Electronic Fingerprints
- f. The Contractor personnel shall submit all required information related to their background investigations (completion of the investigation documents (SF85, SF85P, or SF 86) utilizing the Office of Personnel Management's (OPM) Electronic Questionnaire for Investigations Processing (e-QIP) after receiving an email notification from the Security and Investigation Center (SIC).

Advanced Environmental Controls for Veterans with Severe Disabilities

- g. The Contractor employee shall certify and release the e-QIP document, print and sign the signature pages, and send them encrypted to the COR for electronic submission to the SIC. These documents shall be submitted to the COR within 3 business days of receipt of the e-QIP notification email. (Note: OPM is moving towards a “click to sign” process. If click to sign is used, the Contractor employee should notify the COR within 3 business days that documents were signed via eQIP).
- h. The Contractor shall be responsible for the actions of all personnel provided to work for VA under this contract. In the event that damages arise from work performed by Contractor provided personnel, under the auspices of this contract, the Contractor shall be responsible for all resources necessary to remedy the incident.
- i. A Contractor may be granted unescorted access to VA facilities and/or access to VA Information Technology resources (network and/or protected data) with a favorably adjudicated Special Agreement Check (SAC) or “Closed, No Issues” (SAC) finger print results, training delineated in VA Handbook 6500.6 (Appendix C, Section 9), and, the signed “Contractor Rules of Behavior.” However, the Contractor will be responsible for the actions of the Contractor personnel they provide to perform work for VA. The investigative history for Contractor personnel working under this contract must be maintained in the database of the Office of Personnel Management (OPM).
- j. The Contractor, when notified of an unfavorably adjudicated background investigation on a Contractor employee as determined by the Government, shall withdraw the employee from consideration in working under the contract.
- k. Failure to comply with the Contractor personnel security investigative requirements may result in loss of physical and/or logical access to VA facilities and systems by Contractor and Subcontractor employees and/or termination of the contract for default.
- l. Identity Credential Holders must follow all HSPD-12 policies and procedures as well as use and protect their assigned identity credentials in accordance with VA policies and procedures, displaying their badges at all times, and returning the identity credentials upon termination of their relationship with VA.
- m. Tuberculosis Screening and Testing -
 - 1. For those with previous documented positive purified protein derivative (PPD) test results: All contractor personnel shall provide a note from their physician, dated within the past three months stating they are free of any signs and symptoms of tuberculosis. This evaluation will be renewed annually.
 - 2. For those with previously negative PPD skin test results: personnel shall provide proof of a negative reaction to PPD testing, performed in accordance with the latest CDC standards and State guidelines, with the past six (6) months. This test shall be renewed annually.
 - 3. For PPD skin converters (a change from a previously negative skin test to positive; defined as an increase in duration of 10mm or more within 2

**Advanced Environmental Controls for Veterans
with Severe Disabilities**

years): An evaluation from their physician within 30 days of the positive test stating they are free from signs/symptoms of TB and indicating whether prophylactic treatment is indicated. If treatment is indicated, a note from the physician stating it was satisfactorily completed.

Deliverable:

- A. Contractor Staff Roster

6.4 METHOD AND DISTRIBUTION OF DELIVERABLES

The Contractor shall deliver documentation in electronic format, unless otherwise directed in Section B of the solicitation/contract. Acceptable electronic media include: MS Word 2000/2003/2007/2010, MS Excel 2000/2003/2007/2010, MS PowerPoint 2000/2003/2007/2010, MS Project 2000/2003/2007/2010, MS Access 2000/2003/2007/2010, MS Visio 2000/2002/2003/2007/2010, AutoCAD 2002/2004/2007/2010, and Adobe Postscript Data Format (PDF).

6.5 PERFORMANCE METRICS

The table below defines the Performance Standards and Acceptable Levels of Performance associated with this effort.

Performance Objective	Performance Standard	Acceptable Levels of Performance
A. Technical / Quality of Product or Service	<ol style="list-style-type: none"> 1. Demonstrates understanding of requirements 2. Efficient and effective in meeting requirements 3. Meets technical needs and mission requirements 4. Provides quality services/products 	Satisfactory or higher
B. Project Milestones and Schedule	<ol style="list-style-type: none"> 1. Established milestones and project dates are met 2. Products completed, reviewed, delivered in accordance with the established schedule 3. Notifies customer in advance of potential problems 	Satisfactory or higher

**Advanced Environmental Controls for Veterans
with Severe Disabilities**

C. Cost & Staffing	<ol style="list-style-type: none"> 1. Currency of expertise and staffing levels appropriate 2. Personnel possess necessary knowledge, skills and abilities to perform tasks 	Satisfactory or higher
D. Management	<ol style="list-style-type: none"> 1. Integration and coordination of all activities to execute effort 	Satisfactory or higher

The COR will utilize a Quality Assurance Surveillance Plan (QASP) throughout the life of the contract to ensure that the Contractor is performing the services required by this PWS in an acceptable level of performance. The Government reserves the right to alter or change the surveillance methods in the QASP at its own discretion. A Performance Based Service Assessment will be used by the COR in accordance with the QASP to assess Contractor performance.

6.6 FACILITY/RESOURCE PROVISIONS

The Contractor shall request other Government documentation deemed pertinent to the work accomplishment directly from the Government officials with whom the Contractor has contact. The Contractor shall consider the COR as the final source for needed Government documentation when the Contractor fails to secure the documents by other means. The Contractor is expected to use common knowledge and resourcefulness in securing all other reference materials, standard industry publications, and related materials that are pertinent to the work.

6.7 GOVERNMENT FURNISHED PROPERTY

N/A

6.8 SHIPMENT OF HARDWARE OR EQUIPMENT

N/A

Advanced Environmental Controls for Veterans with Severe Disabilities

ADDENDUM A – ADDITIONAL VA REQUIREMENTS, CONSOLIDATED

A1.0 Cyber and Information Security Requirements for VA IT Services

The Contractor shall ensure adequate LAN/Internet, data, information, and system security in accordance with VA standard operating procedures and standard PWS language, conditions, laws, and regulations. The Contractor's firewall and web server shall meet or exceed VA minimum requirements for security. All VA data shall be protected behind an approved firewall. Any security violations or attempted violations shall be reported to the VA Program Manager and VA Information Security Officer as soon as possible. The Contractor shall follow all applicable VA policies and procedures governing information security, especially those that pertain to certification and accreditation.

Contractor supplied equipment, PCs of all types, equipment with hard drives, etc. for contract services must meet all security requirements that apply to Government Furnished Equipment (GFE) and Government Owned Equipment (GOE). Security Requirements include: a) VA Approved Encryption Software must be installed on all laptops or mobile devices before placed into operation, b) Bluetooth equipped devices are prohibited within VA; Bluetooth must be permanently disabled or removed from the device, c) VA approved anti-virus and firewall software, d) Equipment must meet all VA sanitization requirements and procedures before disposal. The COR, CO, the PM, and the Information Security Officer (ISO) must be notified and verify all security requirements have been adhered to.

Each documented initiative under this contract incorporates VA Handbook 6500.6, "Contract Security," March 12, 2010 by reference as though fully set forth therein. The VA Handbook 6500.6, "Contract Security" shall also be included in every related agreement, contract or order. The VA Handbook 6500.6, Appendix C, is included in this document as Addendum B.

Training requirements: The Contractor shall complete all mandatory training courses on the current VA training site, the VA Talent Management System (TMS), and will be tracked therein. The TMS may be accessed at <https://www.tms.va.gov>. If you do not have a TMS profile, go to <https://www.tms.va.gov> and click on the "Create New User" link on the TMS to gain access.

Contractor employees shall complete a VA Systems Access Agreement if they are provided access privileges as an authorized user of the computer system of VA.

A2.0 VA Enterprise Architecture Compliance

The applications, supplies, and services furnished under this contract must comply with One-VA Enterprise Architecture (EA), available at <http://www.ea.oit.va.gov/index.asp> in force at the time of issuance of this contract, including the Program Management Plan and VA's rules, standards, and guidelines in the Technical Reference Model/Standards Profile (TRMSP). VA reserves the right to assess contract deliverables for EA compliance prior to acceptance.

Advanced Environmental Controls for Veterans with Severe Disabilities

A2.1. VA Internet and Intranet Standards

The Contractor shall adhere to and comply with VA Directive 6102 and VA Handbook 6102, Internet/Intranet Services, including applicable amendments and changes, if the Contractor's work includes managing, maintaining, establishing and presenting information on VA's Internet/Intranet Service Sites. This pertains, but is not limited to: creating announcements; collecting information; databases to be accessed, graphics and links to external sites.

Internet/Intranet Services Directive 6102 is posted at (copy and paste the following URL to browser): http://www1.va.gov/vapubs/viewPublication.asp?Pub_ID=409&FTYPE=2

Internet/Intranet Services Handbook 6102 is posted at (copy and paste following URL to browser): http://www1.va.gov/vapubs/viewPublication.asp?Pub_ID=410&FTYPE=2

A3.0 Notice of the Federal Accessibility Law Affecting All Electronic and Information Technology Procurements (Section 508)

On August 7, 1998, Section 508 of the Rehabilitation Act of 1973 was amended to require that when Federal departments or agencies develop, procure, maintain, or use Electronic and Information Technology, that they shall ensure it allows Federal employees with disabilities to have access to and use of information and data that is comparable to the access to and use of information and data by other Federal employees. Section 508 required the Architectural and Transportation Barriers Compliance Board (Access Board) to publish standards setting forth a definition of electronic and information technology and the technical and functional criteria for such technology to comply with Section 508. These standards have been developed and published with an effective date of December 21, 2000. Federal departments and agencies shall develop all Electronic and Information Technology requirements to comply with the standards found in 36 CFR 1194.

A3.1. Section 508 – Electronic and Information Technology (EIT) Standards

The Section 508 standards established by the Architectural and Transportation Barriers Compliance Board (Access Board) are incorporated into, and made part of all VA orders, solicitations and purchase orders developed to procure Electronic and Information Technology (EIT). These standards are found in their entirety at: <https://www.access-board.gov/guidelines-and-standards/communications-and-it/about-the-section-508-standards/section-508-standards> and <http://www.section508.gov/content/learn/standards>. A printed copy of the standards will be supplied upon request. The Contractor shall comply with the technical standards as marked:

Advanced Environmental Controls for Veterans with Severe Disabilities

- § 1194.21 Software applications and operating systems
- § 1194.22 Web-based intranet and internet information and applications
- § 1194.23 Telecommunications products
- § 1194.24 Video and multimedia products
- § 1194.25 Self contained, closed products
- § 1194.26 Desktop and portable computers
- § 1194.31 Functional Performance Criteria
- § 1194.41 Information, Documentation, and Support

A3.2. Equivalent Facilitation

Alternatively, offerors may propose products and services that provide equivalent facilitation, pursuant to Section 508, subpart A, §1194.5. Such offerors will be considered to have provided equivalent facilitation when the proposed deliverables result in substantially equivalent or greater access to and use of information for those with disabilities.

A3.3. Compatibility with Assistive Technology

The Section 508 standards do not require the installation of specific accessibility-related software or the attachment of an assistive technology device. Section 508 requires that the EIT be compatible with such software and devices so that EIT can be accessible to and usable by individuals using assistive technology, including but not limited to screen readers, screen magnifiers, and speech recognition software.

A3.4. Acceptance and Acceptance Testing

Deliverables resulting from this solicitation will be accepted based in part on satisfaction of the identified Section 508 standards' requirements for accessibility and must include final test results demonstrating Section 508 compliance.

Deliverables should meet applicable accessibility requirements and should not adversely affect accessibility features of existing EIT technologies. The Government reserves the right to independently test for Section 508 Compliance before delivery. The Contractor shall be able to demonstrate Section 508 Compliance upon delivery.

Automated test tools and manual techniques are used in the VA Section 508 compliance assessment. Additional information concerning tools and resources can be found at <http://www.section508.va.gov/section508/Resources.asp>.

Deliverables:

- A. Final Section 508 Compliance Test Results

Advanced Environmental Controls for Veterans with Severe Disabilities

A4.0 Physical Security & Safety Requirements:

The Contractor and their personnel shall follow all VA policies, standard operating procedures, applicable laws and regulations while on VA property. Violations of VA regulations and policies may result in citation and disciplinary measures for persons violating the law.

1. The Contractor and their personnel shall wear visible identification at all times while they are on the premises.
2. VA does not provide parking spaces at the work site; the Contractor must obtain parking at the work site if needed. It is the responsibility of the Contractor to park in the appropriate designated parking areas. VA will not invalidate or make reimbursement for parking violations of the Contractor under any conditions.
3. Smoking is prohibited inside/outside any building other than the designated smoking areas.
4. Possession of weapons is prohibited.
5. The Contractor shall obtain all necessary licenses and/or permits required to perform the work, with the exception of software licenses that need to be procured from a Contractor or vendor in accordance with the requirements document. The Contractor shall take all reasonable precautions necessary to protect persons and property from injury or damage during the performance of this contract.

A5.0 Confidentiality and Non-Disclosure

The Contractor shall follow all VA rules and regulations regarding information security to prevent disclosure of sensitive information to unauthorized individuals or organizations.

The Contractor may have access to Protected Health Information (PHI) and Electronic Protected Health Information (EPHI) that is subject to protection under the regulations issued by the Department of Health and Human Services, as mandated by the Health Insurance Portability and Accountability Act of 1996 (HIPAA); 45 CFR Parts 160 and 164, Subparts A and E, the Standards for Privacy of Individually Identifiable Health Information (“Privacy Rule”); and 45 CFR Parts 160 and 164, Subparts A and C, the Security Standard (“Security Rule”). Pursuant to the Privacy and Security Rules, the Contractor must agree in writing to certain mandatory provisions regarding the use and disclosure of PHI and EPHI.

1. The Contractor will have access to some privileged and confidential materials of VA. These printed and electronic documents are for internal use only, are not to be copied or released without permission, and remain the sole property of VA. Some of these materials are protected by the Privacy Act of 1974 (revised by PL 93-5791) and Title 38. Unauthorized disclosure of Privacy Act or Title 38 covered materials is a criminal offense.
2. The VA CO will be the sole authorized official to release in writing, any data, draft deliverables, final deliverables, or any other written or printed materials pertaining to this contract. The Contractor shall release no information. Any

Advanced Environmental Controls for Veterans with Severe Disabilities

request for information relating to this contract presented to the Contractor shall be submitted to the VA CO for response.

3. Contractor personnel recognize that in the performance of this effort, Contractor personnel may receive or have access to sensitive information, including information provided on a proprietary basis by carriers, equipment manufacturers and other private or public entities. Contractor personnel agree to safeguard such information and use the information exclusively in the performance of this contract. Contractor shall follow all VA rules and regulations regarding information security to prevent disclosure of sensitive information to unauthorized individuals or organizations as enumerated in this section and elsewhere in this Contract and its subparts and appendices.
4. Contractor shall limit access to the minimum number of personnel necessary for contract performance for all information considered sensitive or proprietary in nature. If the Contractor is uncertain of the sensitivity of any information obtained during the performance this contract, the Contractor has a responsibility to ask the VA CO.
5. Contractor shall train all of their employees involved in the performance of this contract on their roles and responsibilities for proper handling and nondisclosure of sensitive VA or proprietary information. Contractor personnel shall not engage in any other action, venture or employment wherein sensitive information shall be used for the profit of any party other than those furnishing the information. The sensitive information transferred, generated, transmitted, or stored herein is for VA benefit and ownership alone.
6. Contractor shall maintain physical security at all facilities housing the activities performed under this contract, including any Contractor facilities according to VA-approved guidelines and directives. The Contractor shall ensure that security procedures are defined and enforced to ensure all personnel who are provided access to patient data must comply with published procedures to protect the privacy and confidentiality of such information as required by VA.
7. Contractor must adhere to the following:
 - a. The use of "thumb drives" or any other medium for transport of information is expressly prohibited.
 - b. Controlled access to system and security software and documentation.
 - c. Recording, monitoring, and control of passwords and privileges.
 - d. All terminated personnel are denied physical and electronic access to all data, program listings, data processing equipment and systems.
 - e. VA, as well as any Contractor (or Subcontractor) systems used to support development, provide the capability to cancel immediately all access privileges and authorizations upon employee termination.
 - f. Contractor PM and VA PM are informed within twenty-four (24) hours of any employee termination.
 - g. Acquisition sensitive information shall be marked "Acquisition Sensitive" and shall be handled as "For Official Use Only (FOUO)".

Advanced Environmental Controls for Veterans with Severe Disabilities

- h. Contractor does not require access to classified data.
8. Regulatory standard of conduct governs all personnel directly and indirectly involved in procurements. All personnel engaged in procurement and related activities shall conduct business in a manner above reproach and, except as authorized by statute or regulation, with complete impartiality and with preferential treatment for none. The general rule is to strictly avoid any conflict of interest or even the appearance of a conflict of interest in VA/Contractor relationships.
9. VA Form 0752 shall be completed by all Contractor employees working on this contract, and shall be provided to the CO before any work is performed. In the case that Contractor personnel are replaced in the future, their replacements shall complete VA Form 0752 prior to beginning work.

A6.0 INFORMATION TECHNOLOGY USING ENERGY-EFFICIENT PRODUCTS

The Contractor shall comply with Sections 524 and Sections 525 of the Energy Independence and Security Act of 2007; Section 104 of the Energy Policy Act of 2005; Executive Order 13693, "Planning for Federal Sustainability in the Next Decade", dated March 19, 2015; Executive Order 13221, "Energy-Efficient Standby Power Devices," dated August 2, 2001; and the Federal Acquisition Regulation (FAR) to provide ENERGY STAR®, Federal Energy Management Program (FEMP) designated, low standby power, and Electronic Product Environmental Assessment Tool (EPEAT) registered products in providing information technology products and/or services.

The Contractor shall ensure that information technology products are procured and/or services are performed with products that meet and/or exceed ENERGY STAR, FEMP designated, low standby power, and EPEAT guidelines. The Contractor shall provide/use products that earn the ENERGY STAR label and meet the ENERGY STAR specifications for energy efficiency. Specifically, the Contractor shall:

1. Provide/use ENERGY STAR products, as specified at www.energystar.gov/products (contains complete product specifications and updated lists of qualifying products).
2. Provide/use the purchasing specifications listed for FEMP designated products at https://www4.eere.energy.gov/femp/requirements/laws_and_requirements/energy_star_and_femp_designated_products_procurement_requirements. The Contractor shall use the low standby power products specified at <http://energy.gov/eere/femp/low-standby-power-products>.
3. Provide/use EPEAT registered products as specified at www.epeat.net. At a minimum, the Contractor shall acquire EPEAT® Bronze registered products. EPEAT registered products are required to meet the technical specifications of ENERGY STAR, but are not automatically on the ENERGY STAR qualified product lists. The Contractor shall ensure that applicable products are on both the EPEAT Registry and ENERGY STAR Qualified Product Lists.

Advanced Environmental Controls for Veterans with Severe Disabilities

4. The Contractor shall use these products to the maximum extent possible without jeopardizing the intended end use or detracting from the overall quality delivered to the end user.

The following is a list of information technology products for which ENERGY STAR, FEMP designated, low standby power, and EPEAT registered products are available:

1. Computer Desktops, Laptops, Notebooks, Displays, Monitors, Integrated Desktop Computers, Workstation Desktops, Thin Clients, Disk Drives
2. Imaging Equipment (Printers Copiers, Multi-Function Devices, Scanners, Fax Machines, Digital Duplicators, Mailing Machines)
3. Televisions, Multimedia Projectors

This list is continually evolving, and as a result is not all-inclusive.

DRAFT

**Advanced Environmental Controls for Veterans
with Severe Disabilities**

**ADDENDUM B – VA INFORMATION AND INFORMATION SYSTEM
SECURITY/PRIVACY LANGUAGE**

**APPLICABLE PARAGRAPHS TAILORED FROM: *THE VA INFORMATION AND
INFORMATION SYSTEM SECURITY/PRIVACY LANGUAGE, VA HANDBOOK
6500.6, APPENDIX C, MARCH 12, 2010***

B1. GENERAL

Contractors, Contractor personnel, Subcontractors, and Subcontractor personnel shall be subject to the same Federal laws, regulations, standards, and VA Directives and Handbooks as VA and VA personnel regarding information and information system security.

B2. ACCESS TO VA INFORMATION AND VA INFORMATION SYSTEMS

a. A Contractor/Subcontractor shall request logical (technical) or physical access to VA information and VA information systems for their employees, Subcontractors, and affiliates only to the extent necessary to perform the services specified in the contract, agreement, or task order.

b. All Contractors, Subcontractors, and third-party servicers and associates working with VA information are subject to the same investigative requirements as those of VA appointees or employees who have access to the same types of information. The level and process of background security investigations for Contractors must be in accordance with VA Directive and Handbook 0710, *Personnel Suitability and Security Program*. The Office for Operations, Security, and Preparedness is responsible for these policies and procedures.

c. Contract personnel who require access to national security programs must have a valid security clearance. National Industrial Security Program (NISP) was established by Executive Order 12829 to ensure that cleared U.S. defense industry contract personnel safeguard the classified information in their possession while performing work on contracts, programs, bids, or research and development efforts. The Department of Veterans Affairs does not have a Memorandum of Agreement with Defense Security Service (DSS). Verification of a Security Clearance must be processed through the Special Security Officer located in the Planning and National Security Service within the Office of Operations, Security, and Preparedness.

d. Custom software development and outsourced operations must be located in the U.S. to the maximum extent practical. If such services are proposed to be performed abroad and are not disallowed by other VA policy or mandates (e.g. Business Associate Agreement, Section 3G), the Contractor/Subcontractor must state where all non-U.S. services are provided and detail a security plan, deemed to be acceptable by VA, specifically to address mitigation of the resulting problems of communication, control, data protection, and so forth. Location within the U.S. may be an evaluation factor.

Advanced Environmental Controls for Veterans with Severe Disabilities

e. The Contractor or Subcontractor must notify the CO immediately when an employee working on a VA system or with access to VA information is reassigned or leaves the Contractor or Subcontractor's employ. The CO must also be notified immediately by the Contractor or Subcontractor prior to an unfriendly termination.

B3. VA INFORMATION CUSTODIAL LANGUAGE

1. Information made available to the Contractor or Subcontractor by VA for the performance or administration of this contract or information developed by the Contractor/Subcontractor in performance or administration of the contract shall be used only for those purposes and shall not be used in any other way without the prior written agreement of VA. This clause expressly limits the Contractor/Subcontractor's rights to use data as described in Rights in Data - General, FAR 52.227-14(d) (1).

2. VA information should not be co-mingled, if possible, with any other data on the Contractors/Subcontractor's information systems or media storage systems in order to ensure VA requirements related to data protection and media sanitization can be met. If co-mingling must be allowed to meet the requirements of the business need, the Contractor must ensure that VA information is returned to VA or destroyed in accordance with VA's sanitization requirements. VA reserves the right to conduct on site inspections of Contractor and Subcontractor IT resources to ensure data security controls, separation of data and job duties, and destruction/media sanitization procedures are in compliance with VA directive requirements.

3. Prior to termination or completion of this contract, Contractor/Subcontractor must not destroy information received from VA, or gathered/created by the Contractor in the course of performing this contract without prior written approval by VA. Any data destruction done on behalf of VA by a Contractor/Subcontractor must be done in accordance with National Archives and Records Administration (NARA) requirements as outlined in VA Directive 6300, *Records and Information Management* and its Handbook 6300.1 *Records Management Procedures*, applicable VA Records Control Schedules, and VA Handbook 6500.1, *Electronic Media Sanitization*. Self-certification by the Contractor that the data destruction requirements above have been met must be sent to the VA CO within 30 days of termination of the contract.

4. The Contractor/Subcontractor must receive, gather, store, back up, maintain, use, disclose and dispose of VA information only in compliance with the terms of the contract and applicable Federal and VA information confidentiality and security laws, regulations and policies. If Federal or VA information confidentiality and security laws, regulations and policies become applicable to VA information or information systems after execution of the contract, or if NIST issues or updates applicable FIPS or Special Publications (SP) after execution of this contract, the parties agree to negotiate in good faith to implement the information confidentiality and security laws, regulations and policies in this contract.

Advanced Environmental Controls for Veterans with Severe Disabilities

5. The Contractor/Subcontractor shall not make copies of VA information except as authorized and necessary to perform the terms of the agreement or to preserve electronic information stored on Contractor/Subcontractor electronic storage media for restoration in case any electronic equipment or data used by the Contractor/Subcontractor needs to be restored to an operating state. If copies are made for restoration purposes, after the restoration is complete, the copies must be appropriately destroyed.

6. If VA determines that the Contractor has violated any of the information confidentiality, privacy, and security provisions of the contract, it shall be sufficient grounds for VA to withhold payment to the Contractor or third party or terminate the contract for default or terminate for cause under Federal Acquisition Regulation (FAR) part 12.

7. If a VHA contract is terminated for cause, the associated Business Associate Agreement (BAA) must also be terminated and appropriate actions taken in accordance with VHA Handbook 1600.05, *Business Associate Agreements*. Absent an agreement to use or disclose protected health information, there is no business associate relationship.

8. The Contractor/Subcontractor must store, transport, or transmit VA sensitive information in an encrypted form, using VA-approved encryption tools that are, at a minimum, FIPS 140-2 validated.

9. The Contractor/Subcontractor's firewall and Web services security controls, if applicable, shall meet or exceed VA minimum requirements. VA Configuration Guidelines are available upon request.

10. Except for uses and disclosures of VA information authorized by this contract for performance of the contract, the Contractor/Subcontractor may use and disclose VA information only in two other situations: (i) in response to a qualifying order of a court of competent jurisdiction, or (ii) with VA prior written approval. The Contractor/Subcontractor must refer all requests for, demands for production of, or inquiries about, VA information and information systems to the VA CO for response.

11. Notwithstanding the provision above, the Contractor/Subcontractor shall not release VA records protected by Title 38 U.S.C. 5705, confidentiality of medical quality assurance records and/or Title 38 U.S.C. 7332, confidentiality of certain health records pertaining to drug addiction, sickle cell anemia, alcoholism or alcohol abuse, or infection with human immunodeficiency virus. If the Contractor/Subcontractor is in receipt of a court order or other requests for the above mentioned information, that Contractor/Subcontractor shall immediately refer such court orders or other requests to the VA CO for response.

12. For service that involves the storage, generating, transmitting, or exchanging of VA sensitive information but does not require C&A or a Memorandum of Understanding-Interconnection Service Agreement (MOU-ISA) for system interconnection, the

Advanced Environmental Controls for Veterans with Severe Disabilities

Contractor/Subcontractor must complete a Contractor Security Control Assessment (CSCA) on a yearly basis and provide it to the COR.

B4. INFORMATION SYSTEM DESIGN AND DEVELOPMENT

1. Information systems that are designed or developed for or on behalf of VA at non-VA facilities shall comply with all VA directives developed in accordance with FISMA, HIPAA, NIST, and related VA security and privacy control requirements for Federal information systems. This includes standards for the protection of electronic PHI, outlined in 45 C.F.R. Part 164, Subpart C, information and system security categorization level designations in accordance with FIPS 199 and FIPS 200 with implementation of all baseline security controls commensurate with the FIPS 199 system security categorization (reference VA Handbook 6500, *Risk Management Framework for VA Information Systems – Tier 3: VA Information Security Program*, and the TIC Reference Architecture). During the development cycle a Privacy Impact Assessment (PIA) must be completed, provided to the COR, and approved by the VA Privacy Service in accordance with Directive 6508, *Implementation of Privacy Threshold Analysis and Privacy Impact Assessment*.

2. The Contractor/Subcontractor shall certify to the COR that applications are fully functional and operate correctly as intended on systems using the VA Federal Desktop Core Configuration (FDCC), and the common security configuration guidelines provided by NIST or VA. This includes Internet Explorer 7 configured to operate on Windows XP and Vista (in Protected Mode on Vista) and future versions, as required.

3. The standard installation, operation, maintenance, updating, and patching of software shall not alter the configuration settings from the VA approved and FDCC configuration. Information technology staff must also use the Windows Installer Service for installation to the default “program files” directory and silently install and uninstall.

4. Applications designed for normal end users shall run in the standard user context without elevated system administration privileges.

5. The security controls must be designed, developed, approved by VA, and implemented in accordance with the provisions of VA security system development life cycle as outlined in NIST Special Publication 800-37, *Guide for Applying the Risk Management Framework to Federal Information Systems*, VA Handbook 6500, *Risk Management Framework for VA Information Systems – Tier 3: VA Information Security Program* and VA Handbook 6500.5, *Incorporating Security and Privacy in System Development Lifecycle*.

6. The Contractor/Subcontractor is required to design, develop, or operate a System of Records Notice (SOR) on individuals to accomplish an agency function subject to the Privacy Act of 1974, (as amended), Public Law 93-579, December 31,

Advanced Environmental Controls for Veterans with Severe Disabilities

1974 (5 U.S.C. 552a) and applicable agency regulations. Violation of the Privacy Act may involve the imposition of criminal and civil penalties.

7. The Contractor/Subcontractor agrees to:

a. Comply with the Privacy Act of 1974 (the Act) and the agency rules and regulations issued under the Act in the design, development, or operation of any system of records on individuals to accomplish an agency function when the contract specifically identifies:

i. The Systems of Records (SOR); and

ii. The design, development, or operation work that the Contractor/Subcontractor is to perform;

b. Include the Privacy Act notification contained in this contract in every solicitation and resulting subcontract and in every subcontract awarded without a solicitation, when the work statement in the proposed subcontract requires the redesign, development, or operation of a SOR on individuals that is subject to the Privacy Act; and

c. Include this Privacy Act clause, including this subparagraph (3), in all subcontracts awarded under this contract which requires the design, development, or operation of such a SOR.

8. In the event of violations of the Act, a civil action may be brought against the agency involved when the violation concerns the design, development, or operation of a SOR on individuals to accomplish an agency function, and criminal penalties may be imposed upon the officers or employees of the agency when the violation concerns the operation of a SOR on individuals to accomplish an agency function. For purposes of the Act, when the contract is for the operation of a SOR on individuals to accomplish an agency function, the Contractor/Subcontractor is considered to be an employee of the agency.

a. "Operation of a System of Records" means performance of any of the activities associated with maintaining the SOR, including the collection, use, maintenance, and dissemination of records.

b. "Record" means any item, collection, or grouping of information about an individual that is maintained by an agency, including, but not limited to, education, financial transactions, medical history, and criminal or employment history and contains the person's name, or identifying number, symbol, or any other identifying particular assigned to the individual, such as a fingerprint or voiceprint, or a photograph.

c. "System of Records" means a group of any records under the control of any agency from which information is retrieved by the name of the individual or by some identifying number, symbol, or other identifying particular assigned to the individual.

Advanced Environmental Controls for Veterans with Severe Disabilities

9. The vendor shall ensure the security of all procured or developed systems and technologies, including their subcomponents (hereinafter referred to as “Systems”), throughout the life of this contract and any extension, warranty, or maintenance periods. This includes, but is not limited to workarounds, patches, hot fixes, upgrades, and any physical components (hereafter referred to as Security Fixes) which may be necessary to fix all security vulnerabilities published or known to the vendor anywhere in the Systems, including Operating Systems and firmware. The vendor shall ensure that Security Fixes shall not negatively impact the Systems.

10. The vendor shall notify VA within 24 hours of the discovery or disclosure of successful exploits of the vulnerability which can compromise the security of the Systems (including the confidentiality or integrity of its data and operations, or the availability of the system). Such issues shall be remediated as quickly as is practical, but in no event longer than 1 day.

11. When the Security Fixes involve installing third party patches (such as Microsoft OS patches or Adobe Acrobat), the vendor will provide written notice to VA that the patch has been validated as not affecting the Systems within 10 working days. When the vendor is responsible for operations or maintenance of the Systems, they shall apply the Security Fixes within 1 day.

12. All other vulnerabilities shall be remediated as specified in this paragraph in a timely manner based on risk, but within 60 days of discovery or disclosure. Exceptions to this paragraph (e.g. for the convenience of VA) shall only be granted with approval of the CO and the VA Assistant Secretary for Office of Information and Technology.

B5. INFORMATION SYSTEM HOSTING, OPERATION, MAINTENANCE, OR USE

a. For information systems that are hosted, operated, maintained, or used on behalf of VA at non-VA facilities, Contractors/Subcontractors are fully responsible and accountable for ensuring compliance with all HIPAA, Privacy Act, FISMA, NIST, FIPS, and VA security and privacy directives and handbooks. This includes conducting compliant risk assessments, routine vulnerability scanning, system patching and change management procedures, and the completion of an acceptable contingency plan for each system. The Contractor's security control procedures must be equivalent, to those procedures used to secure VA systems. A Privacy Impact Assessment (PIA) must also be provided to the COR and approved by VA Privacy Service prior to operational approval. All external Internet connections to VA network involving VA information must be in accordance with the TIC Reference Architecture and reviewed and approved by VA prior to implementation.

b. Adequate security controls for collecting, processing, transmitting, and storing of Personally Identifiable Information (PII), as determined by the VA Privacy Service, must be in place, tested, and approved by VA prior to hosting, operation, maintenance, or use

Advanced Environmental Controls for Veterans with Severe Disabilities

of the information system, or systems by or on behalf of VA. These security controls are to be assessed and stated within the PIA and if these controls are determined not to be in place, or inadequate, a Plan of Action and Milestones (POA&M) must be submitted and approved prior to the collection of PII.

c. Outsourcing (Contractor facility, Contractor equipment or Contractor staff) of systems or network operations, telecommunications services, or other managed services requires certification and accreditation (authorization) (C&A) of the Contractor's systems in accordance with VA Handbook 6500.3, *Assessment, Authorization and Continuous Monitoring of VA Information Systems* and/or the VA OCS Certification Program Office. Government-owned (Government facility or Government equipment) Contractor-operated systems, third party or business partner networks require memorandums of understanding and interconnection agreements (MOU-ISA) which detail what data types are shared, who has access, and the appropriate level of security controls for all systems connected to VA networks.

d. The Contractor/Subcontractor's system must adhere to all FISMA, FIPS, and NIST standards related to the annual FISMA security controls assessment and review and update the PIA. Any deficiencies noted during this assessment must be provided to the VA CO and the ISO for entry into the VA POA&M management process. The Contractor/Subcontractor must use the VA POA&M process to document planned remedial actions to address any deficiencies in information security policies, procedures, and practices, and the completion of those activities. Security deficiencies must be corrected within the timeframes approved by the Government. Contractor/Subcontractor procedures are subject to periodic, unannounced assessments by VA officials, including the VA Office of Inspector General. The physical security aspects associated with Contractor/Subcontractor activities must also be subject to such assessments. If major changes to the system occur that may affect the privacy or security of the data or the system, the C&A of the system may need to be reviewed, retested and re-authorized per VA Handbook 6500.3. This may require reviewing and updating all of the documentation (PIA, System Security Plan, and Contingency Plan). The Certification Program Office can provide guidance on whether a new C&A would be necessary.

e. The Contractor/Subcontractor must conduct an annual self assessment on all systems and outsourced services as required. Both hard copy and electronic copies of the assessment must be provided to the COR. The Government reserves the right to conduct such an assessment using Government personnel or another Contractor/Subcontractor. The Contractor/Subcontractor must take appropriate and timely action (this can be specified in the contract) to correct or mitigate any weaknesses discovered during such testing, generally at no additional cost.

f. VA prohibits the installation and use of personally-owned or Contractor/Subcontractor owned equipment or software on the VA network. If non-VA owned equipment must be used to fulfill the requirements of a contract, it must be stated in the service agreement, SOW or contract. All of the security controls required

Advanced Environmental Controls for Veterans with Severe Disabilities

for Government furnished equipment (GFE) must be utilized in approved other equipment (OE) and must be funded by the owner of the equipment. All remote systems must be equipped with, and use, a VA-approved antivirus (AV) software and a personal (host-based or enclave based) firewall that is configured with a VA approved configuration. Software must be kept current, including all critical updates and patches. Owners of approved OE are responsible for providing and maintaining the anti-viral software and the firewall on the non-VA owned OE.

g. All electronic storage media used on non-VA leased or non-VA owned IT equipment that is used to store, process, or access VA information must be handled in adherence with VA Handbook 6500.1, *Electronic Media Sanitization* upon: (i) completion or termination of the contract or (ii) disposal or return of the IT equipment by the Contractor/Subcontractor or any person acting on behalf of the Contractor/Subcontractor, whichever is earlier. Media (hard drives, optical disks, CDs, back-up tapes, etc.) used by the Contractors/Subcontractors that contain VA information must be returned to VA for sanitization or destruction or the Contractor/Subcontractor must self-certify that the media has been disposed of per 6500.1 requirements. This must be completed within 30 days of termination of the contract.

h. Bio-Medical devices and other equipment or systems containing media (hard drives, optical disks, etc.) with VA sensitive information must not be returned to the vendor at the end of lease, for trade-in, or other purposes. The options are:

- 1) Vendor must accept the system without the drive;
- 2) VA's initial medical device purchase includes a spare drive which must be installed in place of the original drive at time of turn-in; or
- 3) VA must reimburse the company for media at a reasonable open market replacement cost at time of purchase.
- 4) Due to the highly specialized and sometimes proprietary hardware and software associated with medical equipment/systems, if it is not possible for VA to retain the hard drive, then;
 - a) The equipment vendor must have an existing BAA if the device being traded in has sensitive information stored on it and hard drive(s) from the system are being returned physically intact; and
 - b) Any fixed hard drive on the device must be non-destructively sanitized to the greatest extent possible without negatively impacting system operation. Selective clearing down to patient data folder level is recommended using VA approved and validated overwriting technologies/methods/tools. Applicable media sanitization specifications need to be preapproved and described in the purchase order or contract.

Advanced Environmental Controls for Veterans with Severe Disabilities

c) A statement needs to be signed by the Director (System Owner) that states that the drive could not be removed and that (a) and (b) controls above are in place and completed. The ISO needs to maintain the documentation.

B6. SECURITY INCIDENT INVESTIGATION

a. The term “security incident” means an event that has, or could have, resulted in unauthorized access to, loss or damage to VA assets, or sensitive information, or an action that breaches VA security procedures. The Contractor/Subcontractor shall immediately notify the COR and simultaneously, the designated ISO and Privacy Officer for the contract of any known or suspected security/privacy incidents, or any unauthorized disclosure of sensitive information, including that contained in system(s) to which the Contractor/Subcontractor has access.

b. To the extent known by the Contractor/Subcontractor, the Contractor/Subcontractor’s notice to VA shall identify the information involved, the circumstances surrounding the incident (including to whom, how, when, and where the VA information or assets were placed at risk or compromised), and any other information that the Contractor/Subcontractor considers relevant.

c. With respect to unsecured protected health information, the business associate is deemed to have discovered a data breach when the business associate knew or should have known of a breach of such information. Upon discovery, the business associate must notify the covered entity of the breach. Notifications need to be made in accordance with the executed business associate agreement.

d. In instances of theft or break-in or other criminal activity, the Contractor/Subcontractor must concurrently report the incident to the appropriate law enforcement entity (or entities) of jurisdiction, including the VA OIG and Security and Law Enforcement. The Contractor, its employees, and its Subcontractors and their employees shall cooperate with VA and any law enforcement authority responsible for the investigation and prosecution of any possible criminal law violation(s) associated with any incident. The Contractor/Subcontractor shall cooperate with VA in any civil litigation to recover VA information, obtain monetary or other compensation from a third party for damages arising from any incident, or obtain injunctive relief against any third party arising from, or related to, the incident.

B7. LIQUIDATED DAMAGES FOR DATA BREACH

a. Consistent with the requirements of 38 U.S.C. §5725, a contract may require access to sensitive personal information. If so, the Contractor is liable to VA for liquidated damages in the event of a data breach or privacy incident involving any SPI the Contractor/Subcontractor processes or maintains under this contract. However, it is the policy of VA to forgo collection of liquidated damages in the event the Contractor provides payment of actual damages in an amount determined to be adequate by the agency.

Advanced Environmental Controls for Veterans with Severe Disabilities

b. The Contractor/Subcontractor shall provide notice to VA of a “security incident” as set forth in the Security Incident Investigation section above. Upon such notification, VA must secure from a non-Department entity or the VA Office of Inspector General an independent risk analysis of the data breach to determine the level of risk associated with the data breach for the potential misuse of any sensitive personal information involved in the data breach. The term 'data breach' means the loss, theft, or other unauthorized access, or any access other than that incidental to the scope of employment, to data containing sensitive personal information, in electronic or printed form, that results in the potential compromise of the confidentiality or integrity of the data. Contractor shall fully cooperate with the entity performing the risk analysis. Failure to cooperate may be deemed a material breach and grounds for contract termination.

c. Each risk analysis shall address all relevant information concerning the data breach, including the following:

- 1) Nature of the event (loss, theft, unauthorized access);
- 2) Description of the event, including:
 - a) date of occurrence;
 - b) data elements involved, including any PII, such as full name, social security number, date of birth, home address, account number, disability code;
- 3) Number of individuals affected or potentially affected;
- 4) Names of individuals or groups affected or potentially affected;
- 5) Ease of logical data access to the lost, stolen or improperly accessed data in light of the degree of protection for the data, e.g., unencrypted, plain text;
- 6) Amount of time the data has been out of VA control;
- 7) The likelihood that the sensitive personal information will or has been compromised (made accessible to and usable by unauthorized persons);
- 8) Known misuses of data containing sensitive personal information, if any;
- 9) Assessment of the potential harm to the affected individuals;
- 10) Data breach analysis as outlined in 6500.2 Handbook, *Management of Breaches Involving Sensitive Personal Information*, as appropriate; and
- 11) Whether credit protection services may assist record subjects in avoiding or mitigating the results of identity theft based on the sensitive personal information that may have been compromised.

Advanced Environmental Controls for Veterans with Severe Disabilities

d. Based on the determinations of the independent risk analysis, the Contractor shall be responsible for paying to VA liquidated damages in the amount of \$37.50 per affected individual to cover the cost of providing credit protection services to affected individuals consisting of the following:

- 1) Notification;
- 2) One year of credit monitoring services consisting of automatic daily monitoring of at least 3 relevant credit bureau reports;
- 3) Data breach analysis;
- 4) Fraud resolution services, including writing dispute letters, initiating fraud alerts and credit freezes, to assist affected individuals to bring matters to resolution;
- 5) One year of identity theft insurance with \$20,000.00 coverage at \$0 deductible; and
- 6) Necessary legal expenses the subjects may incur to repair falsified or damaged credit records, histories, or financial affairs.

B8. SECURITY CONTROLS COMPLIANCE TESTING

On a periodic basis, VA, including the Office of Inspector General, reserves the right to evaluate any or all of the security controls and privacy practices implemented by the Contractor under the clauses contained within the contract. With 10 working-day's notice, at the request of the Government, the Contractor must fully cooperate and assist in a Government-sponsored security controls assessment at each location wherein VA information is processed or stored, or information systems are developed, operated, maintained, or used on behalf of VA, including those initiated by the Office of Inspector General. The Government may conduct a security control assessment on shorter notice (to include unannounced assessments) as determined by VA in the event of a security incident or at any other time.

B9. TRAINING

- a. All Contractor employees and Subcontractor employees requiring access to VA information and VA information systems shall complete the following before being granted access to VA information and its systems:
 - 1) Successfully complete the *VA Privacy and Information Security Awareness and Rules of Behavior* course (TMS #10176) and complete this required privacy and security training annually; Sign and acknowledge (electronically through TMS #10176) understanding of and responsibilities for compliance with the *Contractor Rules of Behavior*, Appendix D relating to access to VA information and information systems.
 - 2) Successfully complete any additional cyber security or privacy training, as required for VA personnel with equivalent information system access *[to be defined by the VA program official and provided to the CO for inclusion in the solicitation document – e.g., any role-based information security training required in accordance*

**Advanced Environmental Controls for Veterans
with Severe Disabilities**

with NIST Special Publication 800-16, Information Technology Security Training Requirements.]

- b. The Contractor shall provide to the CO and/or the COR a copy of the training certificates and certification of signing the Contractor Rules of Behavior for each applicable employee within 1 week of the initiation of the contract and annually thereafter, as required.
- c. Failure to complete the mandatory annual training and electronically sign the Rules of Behavior annually, within the timeframe required, is grounds for suspension or termination of all physical or electronic access privileges and removal from work on the contract until such time as the training and documents are complete.

DRAFT